Visit: www.thesafetyequipmentstore.com Or Email: besafe@thesafetyequipmentstore.com for Sales & Service.

142 Dräger-Tubes for short-term measurements

## Chlorobenzene 5/a

Order No. 67 28 76	Ord	er N	lo.	67	28	76
--------------------	-----	------	-----	----	----	----

Application	Range

Standard Measuring Range:	5 to 200 ppm
Number of Strokes n:	10
Time for Measurement:	approx. 3 min.
Standard Deviation:	± 10 to 15 %
Color Change:	blue → yellow

## Ambient Operating Conditions

Temperature:	10 to 40 °C
Absolute Humidity:	3 to 15 mg $\rm H_2O$ / L

## **Reaction Principle**

a)  $C_6H_5CI + Cr^{VI} \rightarrow HCI$ 

b) HCl + bromophenol blue  $\rightarrow$  yellow reaction product

## Cross Sensitivity

Other chlorinated hydrocarbons are indicated as well, but with different sensitivities. Methylene chloride does not affect the indication. Chlorine and hydrogen chloride in the TLV range are absorbed in the pre-layer, but in such concentrations do not affect the indication.

Distributed By: The Safety Equipment Store® www.thesafetyequipmentstore.com Email: besafe@thesafetyequipmentstore.com

