## Perchloroethylene 0.1/a

Order No. 81 01551



Standard Measuring Range : 0.5 to 4 ppm / 0.1 to 1 ppm Number of Strokes (n) : 3

Time for Measurement : app. 3 min

Standard Deviation :  $\pm$  20 to 25 %

Colour Change : light grey —> / app. 9 min

: light grey -> blue

## **Ambient Operating Conditions**

 $\begin{array}{lll} \textbf{Temperature} & \textbf{:} & 15 \text{ to } 30 \text{ }^{\circ}\text{C} \\ \textbf{Absolute Humidity} & \textbf{:} & \text{max. } 30 \text{ mg H}_2\text{O} \text{ / L} \\ \end{array}$ 

## **Reaction Principle**

a) CCl<sub>2</sub>=CCl<sub>2</sub> + MnO<sub>4</sub> - --> Cl<sub>2</sub>
 b) Cl<sub>2</sub> + Diphenylbenzidine --> blue reaction product

## **Cross Sensitivity**

Halogenated hydrocarbons and free halogens are indicated.

Petroleum vapours reduce the perchloroethylene indication if they exceed the following concentrations:

40 ppm at n=9

160 ppm at n=3